

### Criteria - Total Hip Replacement (excluding resurfacing arthroplasty)

Criteria - A* Ratings	3A*	5A*	7A*	10A*	13A*	15A*
Minimum number of centres outside development centre(s)	3	3	3	3	3	3
Minimum number of surgeons outside of development centre(s)	3	3	3	3	3	3
Minimum total cohort	150	250	350	500	500	500
Minimum at risk at benchmark time	150	225	300	400	400	400
Maximum revision rate ‡	3.0%	3.3%	3.6%	4.2%	5.1%	5.7%
Criteria - A Ratings	3A	5A	7A	10A	13A	15A
Minimum number of centres and surgeons	3	3	3	3	3	3
Minimum total cohort	150	250	350	500	500	500
Minimum at risk at benchmark time	72	68	65	58	51	47
Maximum revision rate ‡	5.0%	5.3%	5.6%	6.2%	7.1%	7.7%
‡ The upper 95% confidence interval for KM revision rate (1 - Survival) must be lower than the specified level						
Criteria - B Ratings	3B	5B	7B	10B	13B	15B
Minimum number of centres and surgeons	1	1	1	1	1	1
Minimum total cohort	100	100	100	100	100	100
Minimum at risk at benchmark time	40	40	40	40	40	40
Maximum value of 95% lower confidence limit for revision rate	3.0%	3.3%	3.6%	4.2%	5.1%	5.7%
Pre-Entry A*	Pre-Entry					
Product launched under Beyond Compliance	Product details supplied to ODEP					

**FOR PUBLIC RELEASE**

**Published 8 May 2025**



Orthopaedic Data  
Evaluation Panel